

**Safety Communication**

**رسالة سلامة**

**Infusion Pump software May Exhibit Various Problems**

<b>Device/ Product Description:</b>	Infusion Pump Systems
<b>Brand:</b>	Alaris™ System PC Unit
<b>Affected product:</b>	<b>Model 8000:</b> software versions 9.5 and prior <b>Model 8015:</b> software versions 9.33 and prior
<b>Manufacturer:</b>	BD Diagnostic Systems
<b>Problem:</b>	<p>Manufacturer has identified issues are as follows:</p> <ul style="list-style-type: none"><li>Issue 1: Software errors related to System Error Code 255-XX-XXX</li><li>Issue 2: Delay Options programming</li><li>Issue 3: Low Battery Alarm Failure</li><li>Issue 4: Keep Vein Open (KVO) / End of Infusion alarms priority</li><li>Issue 5: Use Errors related to Custom Concentration programming</li></ul> <p>The identified issues could result in delay to the start of an infusion, an interruption of therapy, sudden stop, or lowering the infusion rate. This may put patient at greatest risk of harm.</p>
<b>Recommendation /Actions:</b>	<ol style="list-style-type: none"><li>1. Review this notice and ensure that affected personnel are aware of the contents.</li><li>2. Contact the authorized representative for required correction.</li></ol> <p>For more information, Please click <a href="#">here</a>.</p> <p>If you think you had a problem with your device or a device your patient uses, please do not hesitate to report the problem to SFDA through: <a href="#">NCMDR</a> <a href="#">Vigilance system</a> 19999 unified call center</p>

**Devices/Products  
photo:**



**Authorized  
Representative  
Details**

AR name:

Becton Dickinson B.V.

Assigned Contact Person:

Saleh Al-Sohibanie

Mobile/Phone:

0544760333

Email:

saleh\_al\_sohibanei@europe.bd.com

