

Safety Communication

رسالة سلامة

May Have Cybersecurity Vulnerabilities

Device/ Product Description:	Various Imaging and Ultrasound Systems	
Manufacturer:	GE Healthcare	
Problem Summary:	<p>Cybersecurity vulnerabilities whereas unprotected transport of credentials and exposure of sensitive system information to an unauthorized control sphere.</p> <p>For more information, please check Here.</p> <p>If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system 19999 unified call center</p>	
	AR name:	General Electric Healthcare Arabia
	Assigned Contact Person:	samer Albawardi
	Mobile/Phone:	+(966) 555402028
	Email:	KSA.RA@ge.com