

Safety Communication

رسالة سلامة

Issues related to the polyester-based polyurethane (PE-PUR) sound abatement foam used in Philips Continuous and Non- Continuous Ventilators

Device/ Product Description:	CPAP and Bi-Level PAP Devices	
Affected product:	All Devices manufactured before 26 April 2021, All serial numbers	
	Continuous Ventilator, Non-life Supporting	DreamStation ASV
		DreamStation ST, AVAPS
		SystemOne ASV4
		C-Series ASV
		C-Series S/T and AVAPS
		OmniLab Advanced+
	Noncontinuous Ventilator	SystemOne (Q-Series)
		DreamStation
		DreamStation Go
		Dorma 400
		Dorma 500
REMstar SE Auto		
Manufacturer:	Philips Respironics	
Problem:	<p>Issues can result in serious injury which can be life threatening, cause permanent impairment, and/or require medical intervention to preclude permanent impairment:</p> <ol style="list-style-type: none"> 1) polyester-based polyurethane (PE-PUR) foam may degrade into particles which may enter the device's air pathway and be ingested or inhaled by the user 2) The PE-PUR foam may off-gas certain chemicals. The foam degradation may be exacerbated by use of unapproved cleaning methods, such as ozone, and off-gassing may occur during operation. 	

<p>Recommendation /Actions:</p>	<ul style="list-style-type: none"> • Discontinue use of your device and work with your physician or Durable Medical Equipment (DME) provider to determine the most appropriate options for continued treatment. • To continue use of your device due to lack of alternatives, consult with your physician to determine if the benefit of continuing therapy with your device outweighs the risks. • Follow the instructions as indicated in the attached Field Safety Corrective Action. • You can refer to below link for questions and answers: https://www.philips.sa/en/healthcare/e/sleep/communications/src-update • Contact the authorized representative for required corrective action. <p>For more information, please click here.</p> <p>If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system (19999)unified call center</p>	
<p>Devices/Products photo:</p>	<p>Devices/Products Photos are available at the following link: https://www.philips.sa/en/healthcare/e/sleep/communications/src-update</p>	
<p>Authorized Representative Details</p>	<p>AR name:</p>	<p>Philips Healthcare Saudi Arabia Ltd.</p>
	<p>Assigned Contact Person:</p>	<p>Mohammed Alsamhan</p>
	<p>Mobile/Phone:</p>	<p>+966 503879145</p>
	<p>Email:</p>	<p>SFDA.SA.MET@Philips.com KSA.CCC@Philips.com</p>